

Certificate of analysis

Product Name/Description: LA PCR™ Genome DNA Set
Cat. #: 9060
Lot #: Specified on product label
Storage Conditions: 4 degrees C and -20 degrees C
Expiration Date: Specified on product label or package

Package Size: 20 Rxns each

Package Contents:	1. human genome DNA	500 ng/μl	20 μl
	2. E. coli genome DNA	100 ng/μl	20 μl
	3. human-1 primer	10 pmol/μl	20 μl
	4. human-2 primer	10 pmol/μl	20 μl
	5. E. coli-1 primer	10 pmol/μl	20 μl
	6. E. coli-2 primer	10 pmol/μl	20 μl

Note: Storage Condition of component "human genome DNA" and "E. coli genome DNA": Store at 4°C. It can be stored at -20°C. Avoid freeze-thaw cycles.

Quality Control Data:

- 1) A260/A280 of human genome DNA and E. coli genome DNA was >1.8.
- 2) The concentration of human genomic DNA and E. coli genomic DNA were approximately 500 ng/μl and 100 ng/μl, respectively, as judged from the A260 value.
- 3) When PCR was performed using TaKaRa LA Taq (Cat.# RR002A), fragments of 17.5 kbp and 20 kbp were obtained from human genomic DNA and E. coli genomic DNA, respectively.

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department of Takara Bio Inc.

Notice To Purchaser:

Please refer to the website.

Product Documents, Patent Markings, Trademarks, & Safety Information:

Refer to website for availability to download or contact your local office:
www.takara-bio.co.jp (Japanese)
www.takarabio.com (English)

Licence Statements:

Please visit our website and click the "license" button on the product's web page.

©2019 Takara Bio Inc. All Rights Reserved.

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.