

# Certificate of analysis

---

**Product Name/Description:** TaKaRa Mycoplasma qPCR Detection Kit  
**Cat. #:** RR277A  
**Lot #:** Specified on product label  
**Storage Conditions:** -20 degrees C  
**Expiration Date:** Specified on product label or package

**Package Size:** 96 Tests

<b>Package Contents:</b>	1. Probe qPCR Mix	2.5 X	960 µl	x 3
	2. Primer/Probe Mix (Myco)	12.5 X	384 µl	
	3. Primer/Probe Mix (Spike-in)	12.5 X	192 µl	
	4. H2O		1 ml	x 2
	5. Positive Control (Myco)		300 µl	
	6. PC Dilution Buffer		1 ml	x 2
	7. Spike-in Control DNA		1 ml	

**Quality Control Data:**

Real-time PCR test:

Real-time PCR test was performed according to the user manual. Expected amplification was observed by real-time PCR in FAM channel using Positive Control (Myco) as a template. When H2O was used instead of Positive Control (Myco), no specific amplification was observed in FAM channel.

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department of Takara Bio Inc.

**Notice To Purchaser:**

Please refer to the website.

**Product Documents, Patent Markings, Trademarks, & Safety Information:**

Refer to website for availability to download or contact your local office:  
[www.takara-bio.co.jp](http://www.takara-bio.co.jp) (Japanese)  
[www.takarabio.com](http://www.takarabio.com) (English)

**Licence Statements:**

Please visit our website and click the "license" button on the product's web page.

**©2019 Takara Bio Inc. All Rights Reserved.**

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.