

Reagent test – China CDC primer set

Template: Synthesized RNA 5.00E+00–5.00E+04 copies, NTC n = 2
 Primer/probe: China CDC set
 Instrument: Thermal Cycler Dice™ RealTime System III (Takara)

Gene	Init. Qty	Ct (Cp)				
		Cat. # RR064	Cat. # RR600	Cat. # RR601	Cat. # RR650	Cat. # RR651
ORF1ab	NTC	--	--	--	--	--
	5.00E+00	--	--	--	--	--
		39.80	--	--	--	--
	5.00E+01	38.36	--	38.94	39.41	--
		37.47	--	--	--	--
	5.00E+02	35.03	36.96	35.42	38.41	38.50
		34.72	37.32	36.00	37.80	38.06
5.00E+03	31.68	33.36	32.29	34.43	35.37	
5.00E+04	28.59	29.64	29.24	31.40	32.05	
N	NTC	--	--	--	--	--
	5.00E+00	--	--	39.81	--	38.75
		--	--	--	38.23	39.48
	5.00E+01	38.95	39.38	38.82	37.81	37.42
		38.82	39.75	39.85	36.92	37.45
	5.00E+02	35.24	36.50	36.85	34.51	34.25
		36.25	36.36	37.80	34.36	34.73
5.00E+03	33.00	32.60	32.66	31.14	31.33	
5.00E+04	29.52	29.13	28.87	28.04	27.82	

Detection of SARS-CoV-2 using two different targets (ORF1ab and N) with our one-step RT-qPCR kits. These data were generated by Takara Bio using publicly available protocols and the validated primer set from the China CDC. Use of our products in these protocols has not been validated by external agencies nor optimized by our R&D team. These data are provided for illustrative purposes only. Please refer to published literature for more information about these protocols.

Takara Bio USA, Inc.
 United States/Canada: Inside_Sales@takarabio.com or +800.662.2566 • Asia Pacific: +1.650.919.7300 • Europe: +33.(0)1.3904.6880 • Japan: +81.(0)77.565.6999

For Research Use Only. Not for use in diagnostic procedures.

© 2020 Takara Bio Inc. All Rights Reserved. All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions. Additional product, intellectual property, and restricted use information is available at takarabio.com.

