

I. Introduction

5X PrimePath™ Probe qPCR Kit, No ROX, GPR (Cat No. 638347 & 638349) enables the user to perform accurate qPCR on DNA samples. The highly concentrated master mix allows for flexibility in handling larger sample volumes. The entire protocol from setup to detection takes less than one hour, enabling fast results.

II. Required Materials

This protocol applies to the following Takara Bio products:

- 5X PrimePath Probe qPCR Kit, No ROX, GPR (Cat No. 638347 & 638349)

Additional materials required:

- Primers and probes
- Micropipette and tips (with hydrophobic filters)
- Vortex mixer
- Benchtop centrifuge for tubes or plates
- 1.5 ml Eppendorf tubes, 200 µl PCR tubes, or 200 µl PCR plates for sample preparation
- Tubes or plates for real-time PCR
- A real-time PCR machine (real-time systems that do not require ROX reference dye, such as CFX96 Real-Time PCR Detection System, Bio-Rad, etc.)

III. Protocol

1. Mix the tube of 5X PrimePath Probe qPCR Mix by inverting several times then spin down for 10 sec; the mix may appear cloudy after storage, but this does not affect its performance.
2. On ice and using the table below, combine everything for the qPCR reaction mix EXCEPT the DNA sample for all planned reactions, plus 10% of the total reaction mix volume for overage.

NOTES:

- If the mixture is used immediately, it can be combined at room temperature (no ice).
- Do not add the DNA sample at this time. It is listed in the table to indicate what the final mixture will contain.

qPCR reaction mix (per 25 µl[†] reaction)

Reagent	Final concentration	Singleplex	Multiplex (N [†] targets)
5X PrimePath Probe qPCR Mix, No ROX, GPR	1X	5 µl	5 µl
PCR Forward Primer (10 µM)	0.2 µM	0.5 µl	N [†] x 0.5 µl
PCR Reverse Primer (10 µM)	0.2 µM	0.5 µl	N [†] x 0.5 µl
Probe (10 µM)	0.2 µM	0.5 µl	N [†] x 0.5 µl
DNA sample [‡]	—	2 µl [§]	2 µl [§]
RNase Free H ₂ O	—	Up to 25 µl ^{**}	Up to 25 µl ^{**}
Total volume per reaction		25 µl	25 µl

*Mix, as written, can also be used per 20 µl reaction.

†Where N represents the total number of targets of interest. The total reaction volume will still be 25 µl, while 0.5 µl of each primer and probe will be added per target.

‡Do not add to the reaction mix; this will be added to the mix in Step 4.

§Most common sample volume of Takara Bio PCR products. It is possible to add more; however, depending on the chemicals contained in the sample solution, higher sample volume may have adverse effects.

**Adjust the reaction volume according to the recommendations for the real-time PCR instrument used.

5X PrimePath™ Probe qPCR Kit, No ROX, GPR Protocol-At-A-Glance

3. Add [25 µl – Volume of DNA Sample] of the qPCR reaction mixture into a PCR tube or a 96-well PCR plate.
4. Dispense the volume of DNA sample into the PCR tube or plate well. Spin down for 10 sec.
5. Set up the thermal cycler and run the assay using the reaction conditions below:

95°C	20 sec	
40 cycles:		
95°C	1 sec]
60°C	20 sec	

NOTE: Please follow the instruction manual of the real-time qPCR machine used. If the default setting does not work, perform manual analysis per the instruction manual.

6. After the reaction is complete, check the amplification curve. Confirm that the analytical parameter is correct and that the Ct value has been calculated.

Contact Us	
Customer Service/Ordering	Technical Support
tel: 800.662.2566 (toll-free)	tel: 800.662.2566 (toll-free)
fax: 800.424.1350 (toll-free)	fax: 800.424.1350 (toll-free)
web: takarabio.com/service	web: takarabio.com/support
e-mail: ordersUS@takarabio.com	e-mail: technical_support@takarabio.com

Notice to Purchaser

This product is a general purpose reagent intended **For Laboratory Use**. Outside of the United States, this product is intended for research use only unless otherwise stated. This product is not intended for a specific application or made for any particular clinical use. The performance characteristics of this product have not been fully established. It is the user's responsibility to validate the performance of the product, and any component thereof, for any particular use. Resale or transfer of this product, any component thereof, or any substance produced through use of this product, or any component thereof, to any third party is expressly forbidden. To obtain additional rights, please contact your local office or go to our website for additional information.

Your use of this product is also subject to compliance with the licensing requirements, listed below if applicable, and described on the product's web page at <http://www.takarabio.com>. It is your responsibility to review, understand and adhere to any restrictions imposed by these statements.

© 2023 Takara Bio Inc. All Rights Reserved.

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions. Additional product, intellectual property, and restricted use information is available at takarabio.com.

This document has been reviewed and approved by the Quality Department.